

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

MAY 19 2010

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Supervisor/Operator Performing the Verification Procedure:	
DataMaster cdm S/N <u>130666</u>	
Name <u>John Merrion</u>	ID# <u>4785</u> Date <u>5/15/10</u>
A Agency <u>Fairbanks Police Department</u> Phone # <u>450-6500</u>	
Instrument Location <u>911 Cushman Street Fairbanks, Alaska 99701</u>	
B Alco S/N <u>Y301599</u> Target Value <u>.081</u> High Pressure <u>1000</u>	
Alco Test Value Average <u>.081</u>	<u>.081</u>
1 <sup>st</sup> Alco	2 <sup>nd</sup> Alco
Signature <u>[Signature]</u> (OVER)	

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]  
Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this 17th day of June, 2010.

[Signature] (Notary Seal Stamp)  
Carolyn M. Noland  
Notary Public, State of Alaska  
Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130666

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130666

MAY 15, 2010

*CB*  
*5/24/10*

OPERATOR'S NAME:

MERRION/JOHN/E

OPERATOR'S NUMBER: 4785

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI :

VOC

O.L. #: VOC

DEPT/AGENCY: FBK1

CASE/REPORT:

TEST TYPE: V

ALCO TARGET VALUE: .001

ALCO S/N: X301599

--- BREATH ANALYSIS ---

.001 ADJUSTED FOR 29.52 in

ALCO TARGET .079 04:59

BLANK TEST .000 04:59

INTERNAL STANDARD VERIFIED 05:00

ALCO TV 29.52 in .001 05:00

BLANK TEST .000 05:01

SUBJECT SAMPLE *CB* *5/24/10* .000 05:01

BLANK TEST .000 05:02

ALCO TV 29.51 in .001 05:02

BLANK TEST .000 05:03

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130666

MAY 15, 2010

TIME 04:56

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 41c

BAROMETER: 29.52 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()\*+,-./0123456789:;<=>?@ABCDEFGHIJ  
KLMNOPQRSTUVWXYZ[\]^\_`abcde fghijklmno  
pqrstuvwxyz{|}~+